



CO₂ Cuvette for Dräger® ventilators

Mainstream ETCO₂ monitoring



Critical Care ▪ CO₂ Cuvette for Dräger® ventilators

CO₂ Cuvette for Dräger® ventilators

Our CO₂ cuvette for Dräger ventilators is a single use cuvette for mainstream CO₂ monitoring of mechanically ventilated adult patients.

Compliant with all applicable device standards, the cuvette is intended for use with the Dräger Evita Infinity® V500 ICU ventilator and the Dräger Oxylog® 3000 plus transport ventilator.

It should be positioned between the breathing system patient connection port and the catheter mount or patient interface.

Technical specification

| | |
|-----------------------|-----------------------------------|
| ▪ Patient category: | Adult |
| ▪ Material: | ABS |
| ▪ Internal volume: | 6.6 ml |
| ▪ Resistance to flow: | 0.3 cmH ₂ O at 30L/min |
| ▪ Duration of use: | Single use, up to 7 days |



Single use

Simple, economical solution

Mainstream monitoring

Instant ETCO₂ measurement

Low volume

Minimise patient rebreathing

Clinical recommendations: Capnography

AAGBI guidelines recommend continuous capnography in all patients who are anaesthetised or intubated, regardless of their location in the hospital, or the type of airway device used.^[1]

“Failure to use capnography in ventilated patients likely contributed to more than 70% of ICU related deaths. **Increasing use of capnography on ICU is the single change with the greatest potential to prevent deaths**”^[2]

[1] AAGBI Safety Statement, The use of capnography outside the operating theatre: Updated statement from the Association of Anaesthetists of Great Britain & Ireland (AAGBI): May 2011

[2] 4th National Audit Project of The Royal College of Anaesthetists and the Difficult Airway Society: March 2011

| Code | Description | Box Qty. |
|---------|---|----------|
| 2775000 | CO ₂ cuvette for Dräger® ventilators | 25 |

Make an enquiry

*Dräger®, Evita®, Infinity®, and Oxylog® are registered trademarks of Drägerwerk AG & Co. KGaA



Intersurgical Benelux B.V. Vluchtoord 28, 5406 XP, Uden, Nederland
 T: +31 (0)413 243 860 F: +31 (0)413 243 869 info@intersurgical.nl www.intersurgical.nl



The manufacturer Intersurgical Ltd is certified to ISO 9001:2015, ISO 13485:2016 and ISO 14001:2015

Please think before you print
 Save energy and paper.
 If you must print this information sheet please print it double sided.

IS1.51 NL • issue 1 09.19